



MAXIMUM LOAD – 20 KG / 44 LBS
SINGLE STUD INSTALLATION

APPROVED
Fixed Equipment Anchorage
Office of Statewide Health Planning and Development
OPA-0644
on
Tuesday, February 15, 2005
**** Valid for 3 Years Maximum ****
Anthony R. Pike
Anthony R. Pike (916) 654-1477



©COPYRIGHT 2004. THIS DOCUMENT CONTAINS INFORMATION WHICH SHALL NOT BE REPRODUCED, TRANSFERRED TO OTHER DOCUMENTS, DISCLOSED TO OTHERS USED FOR MANUFACTURING, OR ANY OTHER PURPOSE, WITHOUT THE WRITTEN AUTHORIZATION FROM WESTBROOK ENGINEERING, INC.

△ A △ D SEE PAGE 15 △ F △ G SEE PAGE 16 △ I SEE PAGE 17

TABLE A

Component	Size (inches)	Material	Fasteners	Notes
Wall Track △ I	19.0 x 3.5	Westbrook Engineering Proprietary 6063-T6 Aluminum Extrusion	4 ea. 1/4 x 1.75 Phillips Panhead SMS	To a Single 20 Ga 1-1/4 x 3-1/2 or better Steel Stud Minimum 313 lbs pullout Wood or Steel
Slides △ G △ F	2.5 x 7.0 x .5 2.5 x 3.5 x .5	Westbrook Engineering Proprietary 6063-T6 Aluminum Extrusion with Delrin Buffers		
Arms △ A	13.0 x 2.25 x 1.25 18.0 x 2.25 x 1.25	Westbrook Engineering Proprietary 6063-T6 Aluminum Extrusion	3 ea. 5/16-18 x 2 Flathead Torx40TS (each arm)	Arm to Slide
"C" Type Cup Assembly △ D	5.0" diameter	316 Stainless Steel	1/4-28 MS w/Delrin Friction Washer and ESNA nut	Adjusted for cup movement @ 10 Ft-Lbs
Mounting Plate Assembly △ H	5.0" x 8.0" x .125	Westbrook Engineering Proprietary 6063-T6 Aluminum Extrusion	4 ea M5 x 12mm MS and M5 stop nuts or #10-32 x 3/8 PFHMS and press nuts* **D" – 4 ea. #8 or M4 MS (min)**	*Plate to cup **Adapter Plate to Instrument

ULTIMOUNT C WALL MOUNT

Part #: ULT4-13C

Material: See Table A

Tolerance: 0.00 +/- .005
0.000 +/- .003
All dimensions in inches

Drawn by: MM
Date: 8/12/04
File: ULTIMOUNT_C.6

Westbrook Engineering, Inc.
(650) 594-0962